



**icmr**  
INDIAN COUNCIL OF  
MEDICAL RESEARCH  
Serving the nation since 1911

भारतीय आयुर्विज्ञान अनुसंधान परिषद  
स्वास्थ्य अनुसंधान विभाग, स्वास्थ्य और परिवार  
कल्याण मंत्रालय, भारत सरकार

Indian Council of Medical Research  
Department of Health Research, Ministry of Health  
and Family Welfare, Government of India

**Date: 19/05/2020**

## **Revised Guidelines for TrueNat testing for COVID-19**

The TrueNat system is now a comprehensive assay for screening and confirmation of COVID-19 cases.

- 1) The sample is collected in viral lysis buffer and hence biosafety and biosecurity requirements for use of TrueNat machines are minimal.
- 2) Assay: The assay comprises of following two steps.
  - A) **Step 1:** This is E gene screening assay. All samples of suspect COVID-19 should be first tested by this assay. All negatives are to be considered as **true negatives**. All positive samples should be subjected to confirmation by step 2 assay.
  - B) **Step 2:** RdRp gene confirmatory assay. All samples that test positive by this assay must be considered as **true positive**.
- 3) No further RT-PCR based confirmation is required for samples that are positive after step 2 of the assay above.
- 4) All positive and negative results must be reported to ICMR portal in real time manner.